

Accurate USB Data Logger



The factory calibration values were tested by inserting the USB Data Logger into a test chamber and subjected to 5 stable temperature values, which were measured by a reference thermometer, these values were documented as follows;

- 4.992°C
- 10.013°C
- 23.087°C
- 25.122°C
- 30.205°C



The data was subsequently downloaded and analysed using the Microlite software and the results checked against the reference thermometer. You are welcome to install the software and load up the file or peruse the resulting calibration certificate.

Supporting files

- [Microlite software](#)
- [Data File](#)
- [Calibration Certificate](#)
- [Reference instrument calibration certificate](#)